



American
Screening
Corporation

6658 Youree Dr Ste 180 PMB 404
Shreveport, LA 71105
Ph: 318-798-3306 Fax: 318-798-3386
www.americanscreeningcorp.com

INTENDED USE: The Detectabuse controls are designed to monitor and validate the performance of drugs of abuse detection methods at levels established by SAMHSA, CAP/AACC and many state programs. The Detectabuse control urines are compatible with all quantitative and qualitative drug detection procedures which are sufficiently sensitive to detect the control constituents. They should be treated as any "unknown" specimen while following the specific protocol of the assay being used. This product is intended to be used by health care professionals as an integral part of good laboratory practices.

SUMMARY AND EXPLANATION: The DEA exempt Detectabuse product line of controls is manufactured using a human based matrix that has been stabilized to insure that the product will be viable until the date of expiration. Positive controls are spiked with reference drug standards and/or appropriate metabolites that have been obtained from ISO certified manufacturers. Standards are certified by the manufacturer to be at least 98% minimum purity. Specific gravity, pH, and creatinine fall within the limits of normal human urine.

DESCRIPTION: Each bottle contains stabilized human based urine. Positive control urines have been spiked with authentic reference drug standards and/or appropriate metabolites. Negative control urines are certified negative by combination of immunoassay, GC/MS and/or LC/MS for the constituents listed on our target sheets.

STORAGE & STABILITY - Please refer to Limitations for detailed instructions.

Unopened:

A. The controls are stable until the expiration date when stored at -10 to -20 °C and protected from light.

B. The controls are stable until the expiration date when stored at 2-8 °C, however, no stability claims can be made for Oxazepam as it may deteriorate over time when stored refrigerated.

After Opening:

A. The controls are stable for six months or until the expiration date, whichever comes first, when stored at -10 to -20 °C.

(Controls can be aliquoted and frozen)

B. The controls are stable for 31 days or until the expiration date, whichever comes first, when stored tightly capped at 2-8 °C.

C. Thaw controls as needed; allow to come to room temperature followed by gentle swirling before use.

PROCEDURE: Allow controls to come to room temperature followed by gentle swirling or inversion before use. DO NOT SHAKE. Pipette an appropriate aliquot of Detectabuse control urine as required by the drugs of abuse test device or screening method

EXPECTED RESULTS: The positive Detectabuse control must test positive on the drugs of abuse test device or screening method. The negative control must test negative. Biochemical Diagnostics will (upon request), supply assay values derived from our contract assay laboratories and customer base on a particular lot of control material

PRECAUTIONS: For In Vitro Diagnostic Use Only. Please read the entire package insert before using the Detectabuse control urines.

Please use the same safety precautions you would use for processing any "unknown" urine sample containing potentially infectious biological material. Protect product from exposure to direct sunlight. Contains sodium azide: To prevent formation of explosive metal azides dispose of waste by flushing with copious amounts of water or according to local governing regulations.

Do not use beyond the expiration date.

LIMITATIONS OF PROCEDURE: This control is meant to be used to validate the performance of immunoassay drug screening methods.

Consult test manufacturers instructions when using this product; changes in reagents, sample requirement, or methodology may effect test results. Although target values are provided with the Detectabuse liquid controls, each laboratory should run these controls as unknowns in order to establish "in-house" assay values for them. *This product is not meant to be used as a standard or calibrator.*

DETECTABUSE CONTROLS, OXAZEPAM STABILITY: Oxazepam has known stability problems in urine stored refrigerated, and to a lesser degree, frozen. Our experience indicates that Oxazepam will not deteriorate more than 10% of target level for at least one year when stored frozen at -20 °C. Further deteriorations may occur beyond this period although Oxazepam ordinarily tests positive throughout the control's shelf life.

DETECTABUSE CONTROLS, THC STABILITY: Detectabuse controls are stable for the length of time under the storage conditions stated in the package insert. In spite of this fact, under certain conditions, there may be observed a gradual decline in THC levels, over time, from continuous use of a single bottle of control material. This drop in THC values may occur from any THC sample (i.e. calibrators, controls, and samples). The apparent loss of THC most often occurs from handling and not from product instability. It is well known that THC-COOH binds to surfaces, especially certain plastics^{1,2}. In order to minimize this adsorption loss we recommend the following when handling any sample (including Detectabuse controls) which may contain THC: **1.** Preferably, use glass pipettes or pour controls into sample cups. As an alternate, pipettors with disposable plastic tips may be used. Soft plastic transfer pipettes should be avoided. **2.** Do not rinse the pipette back and forth into the sample. **3.** Sample volume to surface area ratio should be as high as possible (i.e. when transferring, sample containers should be filled as much as possible with sample). Avoid rough surface plastic containers. **4.** When pipetting, immerse the pipette tip as little as possible into the sample solution. **5.** Do not return any unused material back into the original sample. These same guidelines should also be followed when aliquoting a control (or sample) for future use.

**DETECTABUSE™ Liquid Control Urine,
Stat-Skreen™ (Low Opiate) or Stat-Skreen-H™ (High Opiate) Target Values (ng/mL)**

SAMHSA MANDATED	-50%	-25%*	Cutoff*	+25%*	+50%	2X	3X
Delta-9-THC-COOH	25	37.5	50	62.5	75	100	150
Benzoyllecgonine	150	225	300	375	450	600	900
Phencyclidine (PCP)	12.5	19	25	31	37.5	50	75
Morphine (Low Opiate)	150	225	300	375	450	600	900
Morphine (High Opiate)	1000	1500	2000	2500	3000	4000	6000
d-Amphetamine	500	750	1000	1250	1500	2000	3000
d-Methamphetamine	500	750	1000	1250	1500	2000	3000
NON-MANDATED							
Secobarbital	150	225	300	375	450	600	900
Oxazepam	150	225	300	375	450	600	900
Methadone	150	225	300	375	450	600	900
Methaqualone	150	225	300	375	450	600	900
Propoxyphene	150	225	300	375	450	600	900
Nortriptyline	500	750	1000	1250	1500	2000	3000
MDMA**	-	-	-	-	-	1000	-
Oxycodone	-	-	-	-	-	200	-
Buprenorphine	-	-	-	-	-	20	-

* Call for Cutoff, Cutoff +25% or Cutoff -25% ordering information

** 3,4 Methylendioxyethyl-amphetamine

CATALOG #	DESCRIPTION	20 mL (High Opiate Controls)
5 mL (High Opiate Controls)		
19470077	Stat-Skreen-H Cutoff -50%	19470071 Stat-Skreen-H Cutoff -50%
19470076	Stat-Skreen-H Cutoff +50%	19470070 Stat-Skreen-H Cutoff +50%
19470081	Stat-Skreen-H 3X Cutoff	19470075 Stat-Skreen-H 3X Cutoff
18001265	ONESCREEN 2X Opiate High Urine Controls	
18002528	ONESCREEN Negative Urine Controls	
5 mL (Low Opiate Controls)		
19470059	Stat-Skreen Cutoff -50%	19470052 Stat-Skreen Cutoff -50%
19470058	Stat-Skreen Cutoff +50%	19470051 Stat-Skreen Cutoff +50%
19470061	Stat-Skreen 3X Cutoff	19470055 Stat-Skreen 3X Cutoff
19700000	Negative*	19223011 Negative*
*Negative Control is Negative for the constituents listed on our target sheets		
19001066	Stat-Skreen with MDMA, OXY & BUP, 2X Cutoff	5 mL (Low Opiate Controls)
19001081	Stat-Skreen-H with MDMA, OXY & BUP, 2X Cutoff	5 mL (High Opiate Controls)
19460007	MDMA 750 ng/mL & Oxycodone 150 ng/mL	5 mL
19470009	Buprenorphine Liquid Control Urine, 15 ng/mL	5 mL
19000102	Cotinine Liquid Control Urine**, 400 ng/mL	5 mL
**Cotinine is a metabolite of Nicotine, used to monitor the performance of Nicotine detection devices		
19470097	Stat-Skreen Multi-Level 6 Set	1 X 20 mL of each level (Low Opiate Controls)
19470099	Stat-Skreen-H Multi-Level 6 Set	1 X 20 mL of each level (High Opiate Controls)
Multi-Level Sets include: Negative, Cutoff -50%, Cutoff +50%, Cutoff -25%, Cutoff +25% and 3X Cutoff		
18000000-18900000	Detectabuse Custom Control Urine***	Bottle Sizes Vary
***Refer to Custom Target Sheets for specific constituents.		